

DENTIS 내부연결형 서브머지드 임플란트에서 지대주 선택에 따른 성공률의 후향적 연구

¹동아대학교의료원 치과보철과, ²동아대학교의료원 구강악안면외과
김은희¹, 이정은¹, 황희성¹, 김철훈², 김정한², 김복주^{2*}

ABSTRACT

A retrospective randomized study of success rates according to abutment selection in DENTIS submerged implant with an internal hex connection

¹Department of Prosthodontics, Dong-a University

²Department of Oral and Maxillofacial Surgery, Implant Clinics, Dong-a University

Eun-Hee Kim¹, Joeng-Eun Lee¹, Bok-Joo Kim^{2*}, Hee-Seong Hwang¹, Chul-Hoon Kim², Jung-Han Kim²

PURPOSE. The purpose of this study is to determine the efficacy of the DENTIS submerged-type implant with an internal hex connection and to build corresponding abutment-selection criteria.

MATERIAL AND METHODS. A total of 204 patients received submerged implant fixtures with an internal hex connection at the Dong-A University Hospital Dental clinic in Busan from January 2013 and May 2016. Three specific abutments, UCLA abutments, customized abutments, ready-made abutments, were randomly selected. Implant success was defined as the basis of the International Congress of Oral Implantologists(ICOI, 2007) criteria. The relationship between the implant success rate and the abutment factor was analyzed using the Kruskal-Wallis test($P<.05$).

RESULTS. A total of 508 implants were placed in 204 patients. After a mean observation period of 38.6 months, 493 out of 508 implants were in normal function, yielding an overall success rate of 97.05%. A total of 15 implants failed: 10 in the maxillary molar area, 4 in the mandibular molar area, and 1 in the mandibular incisal area. All of the implant failures occurred in a single-implant prosthesis, especially high in the maxillary molar area. The Kruskal-Wallis analysis showed that abutment selection has no significant correlation with implant failure($P>.05$).

CONCLUSION. DENTIS submerged implants with an internal hex connection showed predictable results with a success rate of 97.05%. It is no influence on the success rate in the selection of submerged implant abutment with an internal hex connection.

Key words : Dental implants; Survival rate; Dental implant-abutment design;

Corresponding Author

Bok-Joo Kim

Department of Oral and Maxillofacial Surgery, Implant Clinics, Dong-a University, Medical center 26

Daesingongwon-ro, Seo-gu, Busan 49201, Republic of Korea

Tel : 82 51 240 5475, e-mail : omsbjkim@dau.ac.kr

ORCID : 0000-0002-7482-9738

I . Introduction

There were two methods of connecting an implant fixture and an abutment: an external hex connection¹⁾ and an internal hex connection²⁾ (Figs. 1a and b). With an external hex connection, excessive loading at the initial stage could lead to loss of crestal bone and could cause peri-implantitis as a result of bacterial infection arising in the micro-gap between the fixture and the abutment³⁾. Whereas with an internal, tissue-level hex connection, exposure of the metal collar at the thin gingiva could show poor esthetics and make it more difficult to obtain emergency profile. In order to overcome these limitations, the submerged-type implant fixture with an internal hex connection has been designed and developed. In this case, the hexagon is positioned in the fixture, and the fixture platform is located at the alveolar bone level (Fig. 1c)⁴⁾.

Nevertheless, clinicians do not always choose the submerged implant type with an internal hex connection. And they fall into confusion in their choice, either external hex or internal hex. Practitioners lacking in clinical experience could be particularly uncertain of the proper implant fixture to select. After the implant insertion, the selection of the most appropriate superstructure for the prosthesis also could be confusing the dentists.

The abutments introduced for utilization in dental implant superstructures were as follows: casting abutment⁵⁻⁷⁾, which used a metal alloy, customized abutment, which milled the abutment according to its gingival shape and depth of the fixture, and ready-made abutment, which manufacturers produced in fixed sizes (Fig. 2). For the selection of abutment, there are no clear guidelines. Most clinicians rather are obliged to base their judgments only on esthetics, material, the convenience of their treatment, or the



Fig. 1. a. Branemark, external hexagonal connection (Branemark system MKIII, Nobel Biocare AB, Goteborg, Sweden), b. ITI, internal hexagonal connection (Straumann standard implant, Institute Straumann AG, Waldenberg, Switzerland) and c. Submerged internal hexagonal connection (s-Clean tapered II, Dentis, Dae-Gu, Republic of Korea).



Fig. 2. A. Gold UCLA abutment (s-Clean Gold UCLA Abutment, Dentis, Dae-Gu, Republic of Korea), B. Customized abutment (MyPLANT abutment, RaphaBio, Seoul, Republic of Korea), and C. Ready-made abutment (s-Clean Couple Abutment, Dentis, Dae-Gu, Republic of Korea).

practitioner's personal.

Establishment of abutment-selection criteria to help ensure the success of long-term treatment is controversial. For example, the type of gingiva and thickness could be varied in the submerged implant, depending on its located alveolar bone level. This is one of the reasons why abutment selection should be made very carefully.

The purpose of this research is to determine the efficacy of the DENTIS submerged-type implant with an internal hex connection and to build corresponding abutment-selection criteria. To those ends, the comparison of success rates was conducted according to both implant placement and abutment selection.

II. Materials and Methods

Patients who had undergone insertion of DENTIS submerged implant fixture with an internal hex connection between January 2013 and May 2016 at the Dong-A University Hospital Dental Clinic in Busan, Republic of Korea, and who had experienced more than six months of normal function, were selected for inclusion in this study. Patients suffering from early failure prior to prosthesis insertion or from poor oral hygiene after implantation, implant overdenture, or systemic diseases such as uncontrolled osteoporosis or diabetes mellitus, were excluded. The investigation focused on patient data including gender, age, implant placement site, abutment type, and post-prosthetic complication, which information had been obtained from the

relevant charts and radiographic film.

Three specific abutments of the three general types - casting abutment, customized abutment, and ready-made abutment - were adopted for use in this study: UCLA abutment(screw-retained prosthesis/ s-Clean Gold UCLA Abutment, Dentis, Daegu, Republic of Korea), Customized abutment(screw-cement-retained or cement-retained prosthesis/MyPLANT, RaphaBio, Seoul, Republic of Korea), and ready-made abutment(screw-cement-retained or cement-retained prosthesis/s-Clean Couple Abutment, Dentis, Seoul, Republic of Korea), respectively. The abutments used in this study were randomly assigned to the inserted implants. In this research, implant success was defined as the basis of the International Congress of Oral Implantologists(ICOI)[®] criteria(2007): I. Success, and II. Satisfactory survival(Table I). Analysis of statistical difference between the implant success rate and the abutment factor in this study was assessed using the Kruskal-Wallis test($P<.05$).

III. Results

A total of 508 implants of 204 patients (males: 111, females: 93, age: 16-79, average age: 50.1) were selected for this investigation. The time duration from an insertion of the implant fixture to the delivery of prosthesis was an average of 6.6 months. The time duration following prosthesis insertion, which indicated the average maintenance period, was 38.6 months(minimum:

25 months, maximum: 55 months).

A total of 261 implants were inserted into the maxilla: 56 (25 UCLA, 28 MyPLANT, 3 ready-made) into the incisal area: 82 (38 UCLA, 37 MyPLANT, 7 ready-made) into the premolar area, and 123 (75 UCLA, 34 MyPLANT, 14 ready-made) into the molar area (Table II).

A total of 247 implants were inserted into the mandible: 31 (8 UCLA, 16 MyPLANT, 7 ready-made) into the incisal area: 64 (27 UCLA, 30 MyPLANT, 7 ready-made) into the premolar area, and 152 (109 UCLA, 36 MyPLANT, 7 ready-made) into the molar area (Table III).

After prosthesis delivery, 493 implants of 508 were in normal function intraorally, which represented an overall success rate of 97.05%. A

total of 15 implants failed: 10 in the maxillary molar area, 4 in the mandibular molar area, and 1 in the mandibular incisal area. Most of the failures involved the UCLA abutment in the maxillary molar area (Table IV, V).

Seven of the failures occurred within an average of 7 months of prosthesis delivery, 2 within 1 year, and the remaining 6 within 18.5 months. They all were functioning as a single-implant prosthesis.

Regarding the comparison of the three abutments by Kruskal-Wallis test, there was no significant association between implant success rates and the three different abutment types ($P=.420$) (Table VI).

Table 1. The International Congress of Oral Implantologists (ICOI) (2007) classifications: success, survival, failure

Implant Quality Scale Group	Clinical Conditions
I. Success (optimum health)	a) No pain or tenderness upon function b) 0 mobility c) <2mm radiographic bone loss from initial surgery d) No exudates history
II. Satisfactory survival	a) No pain on function b) 0 mobility c) 2-4mm radiographic bone loss d) No exudates history
III. Compromised survival	a) May have sensitivity on function b) No mobility c) Radiographic bone loss >4mm (less than 1/2 of implant body) d) Probing depth >7mm e) May have exudates history
IV. Failure (clinical or absolute failure)	Any of following: a) Pain on function b) Mobility c) Radiographic bone loss >1/2 length of implant d) Uncontrolled exudate e) No longer in mouth

Table II. Abutment-type selection according to the maxillary area

	Incisor(n = 56)	Premolar(n = 82)	Molar(n = 123)
UCLA abutment(n = 138)	25(44.6%)	38(46.3%)	75(61.0%)
MyPLANT abutment(n = 99)	28(50.0%)	37(45.1%)	34(27.6%)
Ready-made abutment(n = 24)	3(5.36%)	7(8.34%)	14(11.4%)




Table III. Abutment-type selection according to the mandibular area

	Incisor(n = 31)	Premolar(n = 64)	Molar(n = 152)
UCLA abutment(n = 144)	8(25.8%)	27(42.2%)	109(71.7%)
MyPLANT abutment(n = 82)	16(51.6%)	30(46.9%)	36(23.7%)
Ready-made abutment(n = 21)	7(22.6%)	7(10.9%)	7(4.61%)




Table IV. Implantation failure according to the abutment selection

	UCLA abutment	MyPLANT abutment	Ready-made abutment
Failure No. / Total No. (Failure No. on the insertion sites)	10 / 282 (Mn. incisor: 1, Mx. / Mn. molar: 7 / 2)	5 / 181 (Mx. / Mn. molar: 3 / 2)	0 / 45
Failure rate(%)	3.54%	2.7%	-

Table V. Implantation failure according to the implantation area

	Incisor	Premolar	Maxilla molar	Mandible molar
Failure No.	1	0	10	4

Table VI. Comparison of three abutments using Kruskal-Wallis test

	UCLA abutment	MyPLANT abutment	Ready-made abutment
N	282	181	45
median	256.01	254.02	247.00
Kruskal-Wallis Test, H Value		1.735	
P-value		.420	

IV. Discussions

When Branemark introduced implantation to dentistry, connection to the abutment was achieved by means of an external hex implant fixture. Adell et al¹⁾. estimated that alveolar bone resorption had advanced an average of 1.2 mm over the course of 1 year following implant fixture insertion and prosthesis delivery. As a resorption prevention, the concept of platform switching was introduced. Becker et al⁹⁾. demonstrated that platform switching might increase the distance between the abutment margin and the alveolar crest, thus decreasing its bone-resorptive effect.

Balfour et al.¹⁰⁾ reported, concerning the evaluation of torsional loading and compressive bending, that increased force is necessary in order to remove a single abutment under internal as opposed to external hex connection. Chang et al¹¹⁾. performed a three-dimensional finite element analysis, the results of which revealed that implants with an external hex connection

were subjected to greater stress than submerged implants with an internal hex connection.

Khraisat et al¹²⁾. and Steinebrunner et al¹³⁾. reported in vitro research findings indicating that internal connection was superior to the external connection in terms of resistance against fatigue, fracture strength, and the failure mode. This suggested that submerged implant with an internal hex connection could be more effective both in the anterior area, where there was significant lateral loading and in the posterior area, where there was substantial occlusal force.

In this study, 508 DENTIS implants were placed in 204 patients, and an overall implantation success rate was 97.05% at 38.6month follow up period. A total of 15 implants failed:10 in the maxillary molar area, 4 in the mandibular molar area, and 1 in the mandibular incisal area. This result could be compared favorably with other cases, where a 5-year implant survival rate under an external hex connection was 84~92% for the maxilla and 91~99% for the mandible¹⁴⁾. Furthermore, the result in this study compared

favorably also within which the 5-year implant survival rate under an internal hex connection did not exceed 90%¹⁵).

All failed implants were functioning as a single-implant prosthesis in this study. Isidor¹⁶ emphasized, based on research on monkeys, that excessive occlusal loading could aggravate loss of osseointegration and occur peri-implantitis. Rangert et al¹⁷. reported that single- or two-implant arrangement in posterior teeth increased the risk of bending overload, resultantly, all single implants in the first-molar area fractured. Goodacre et al¹⁸. conducted a literature review and found that 12 of 332 single implants had to be removed, the majority of which failures were post-prosthetic. Among the other prosthesis types (i.e., implant overdentures, implant fixed partial/complete dentures), pre-prosthetic failure was the most.

There are other important factors that affect the success of the implant. For example, patient's age, gender, systemic condition, implant size, additional surgery, opposing dentition, smoking and splinting were not considered in this study because our study focused on randomly comparing the success rates of implant abutment types over a period of time. In addition, studies of various local and systemic factors affecting the success rate of implant have already proved a number of results, so we assumed that including the evaluation of these factors would cause confusion in our results. Despite of high implant success rates in this research, implant success rate influenced by local and systemic risk factors such as osteoporosis, Crohn's disease, smoking

habits, implant (length, diameter and location) and vicinity with the natural dentition, are still inevitable and showed contradictory results¹⁹). Moy et al²⁰. reported that patients who were over age 60, smoked, had a history of diabetes or radiation treatment, or hormone therapy significantly increased implant failure compared with healthy patients. Chrcanvic et al²¹. assessed the influence of local and systemic factors on the occurrence of dental implant failures up to the second-stage surgery (abutment connection). The distribution of implants in sites of different bone quantities and qualities was quite similar between implants lost up to and after abutment connection. Therefore, relevant factors affecting the implant success rate should be considered with caution.

For use of the submerged implant with an internal hex connection, the abutments selected were casting abutment, customized abutment, and ready-made abutment. The gold UCLA casting prosthesis offered excellent biocompatibility and fracture resistance even when used in reduced crown-height space. Additionally, it allowed convenient condition for insertion or removal of the prosthesis on account of which ease access for repairing was available in the event of encountering troubles²²). Moreover, as it leaves no excess intraoral cement from its retrievability, it could effectively prevent peri-implantitis²³).

However, in cases of UCLA abutment, screw fracture possibly could occur due to screw loosening in a final prosthesis. Indeed, among the disadvantages of the UCLA abutment type were

the complexity and difficulty of prosthesis production, the problematic acquisition of passive fit, and an unaesthetic appearance^{24, 25}. Furthermore, there was the possibility of micro-leakage from screw holes on the occlusal surfaces, which could lead to occlusion problems or poor hygiene if such restorations were not in suitable positions^{26, 27}.

Byrne et al²⁸. emphasized that when using casting abutment, there were, relative to the cases for pre-machined abutment or customized abutment(the shape of which was modified at a laboratory), more serious interfacial and vertical discrepancies. Hebel et al²³. reported the results of tests showing that when non-passive castings were formed for a multi-unit implant prosthesis of the screw-retained type, these misfits between the implant fixture and the abutment left micro gaps.

As confirmed by this research, reduced crown height space, patient preference, the difficulty of porcelain reduction for occlusal adjustment, and repair convenience were the main reasons UCLA was the most commonly used abutment type. The implant failure in this investigation was significantly not affected by abutment design. However, most of the failed implants were those with a single-unit UCLA abutment, in which case the fixing screw made a direct connection between the abutment and the fixture. With this kind of arrangement, there could be a direct occlusal load on the fixture. In looking at implant failure according to the insertion sites, it occurred mostly in the maxillary molar area. It was thought that implant site's bone quality was one of the factors in the implant failure, and this

was supported by the previous research^{29, 30}.

A customized abutment is fabricated by milling it according to its gingival shape and the location of the inserted fixture platform. This could facilitate both the reproducing of an emergency profile and overall laboratory processes. It is important for preventing plaque deposition, ease of oral hygiene and esthetics to make natural emergence profile of prosthesis in the clinical situation. The round ready-made abutment in the occlusal plane did not match an anatomic gingival configuration, therefore customized abutments had efficacy by solving the difference between the ready-made abutment and the cross-sectional form of natural teeth. Especially in the anterior area, where an excellent esthetic was required, irregular gingival level, and deeply inserted site, customized abutments could be useful for an anatomically ideal prosthesis³¹. By contrast, if the ready-made abutment was used in those case, with fixed angulation it might be adjusted excessively, therefore it could lead to reducing the retention of the prosthesis. Moreover, it might result in less than ideal crown contours and insufficient support for optimum soft tissue esthetics because its size of the collar height and the the cross section were fixed³²⁻³⁴. Thus, a customized abutment could make the thickness of the final prosthesis ideal, and enable easy to retrieve splinting prosthesis without loss of retention^{32, 35, 36}. This research indicated that in incisor and premolar placement cases, where esthetics is considered to be important, customized abutment is preferable to the ready-made abutment.

Within the limitation of this study, DENTIS submerged implant with an internal hex connection showed predictable results with a success rate of 97.05% at the maxillary and mandibular incisor, premolar and molar sites. There was no significant association between

implant success rate and the three different abutment types. However, many relative factors affecting implant success rates should be considered, and a long-term follow-up period will be needed to assess implant prognosis.

참 고 문 헌

- Adell R, Lekholm U, Rockler B, Brånemark PI. A 15-year study of osseointegrated implants in the treatment of the edentulous jaw. *Int. J Oral Surg* 1981;10:387-416.
- Maeda Y, Satoh T, Sogo M. In vitro differences of stress concentrations for internal and external hex implant-abutment connections: a short communication. *J Oral Rehab* 2006;33:75-78.
- Oh TJ, Yoon J, Misch CE, Wang HL. The cause of early implant bone loss : myth or science? *J Periodontol* 2002;73:322-333.
- Stephen T, Chen Ivan D, Darby Eric C, Reynolds. A prospective clinical study of non-submerged immediate implants: clinical outcomes and esthetic results. *Clin. Oral Impl. Res* 2007;18:552-562.
- Lewis SG, Beumer J, Perri GR, Hornburg WP. Single tooth implant supported restorations. *Int J Oral Maxillofac Implants* 1988;3:25-30
- Lewis SG, Beumer J, Perri GR, Hornburg WP. The "UCLA" abutment. *Int J Oral Maxillofac Implants* 1988;3:183-189.
- Lewis SG, Llamas D, Avera S. The UCLA abutment: a four-year review. *J Prosthet Dent* 1992;67:509-515.
- Misch CE, Perel ML, Wang HL, Sammartino G, Galindo-Moreno P, Trisi P, Steigmann M, Rebaudi A, Palti A, Pikos MA, Schwartz-Arad D, Choukroun J, Perez JL, Marenzi G, Valavanis DK. Implant success, survival, and failure: The International Congress of Oral Implantologists(ICOI) Pisa Consensus Conference. *Implant Dentistry* 2008;17:5-11.
- Becker J, Ferrari D, Herten M, Kirsch A, Schaer A, Schwarz F. Influence of platform switching on crestal bone changes at non-submerged titanium implants: a histomorphometrical study in dogs. *J Clin Periodontol* 2007; 34:1089-1096.
- Balfour A, O'Brien GR. Comparative study of antirotational single tooth abutments. *J Prosthet Dent* 1995;73:36-43.
- Chang HS, Chen YC, Hsieh YD, Hsu ML. Stress distribution of two commercial dental implant systems: A three-dimensional finite element analysis. *J Dent Sci* 2013;8: 261-271.
- Khraisat A, Stegaroiu R, Nomura S, Miyakawa O. Fatigue resistance of two implant/abutment joint design. *J Prosthet Dent* 2002;88:604-10.
- Steinebrunner L, Wolfart S, Ludwig K, Kern M. Implant-abutment interface design affects fatigue and fracture strength of implants. *Clin Oral Impl Res* 2008;29:1276-1284.
- Adell R, Eriksson B, Lekholm U, Brånemark PI, Jemt T. A long-term follow-up study of osseointegrated implants in the treatment of totally edentulous jaws. *Int J Oral Maxillofac Implants* 1990;5:347-359.
- Buser D, Mericske-Stern R, Bernard JP, Behneke N, Hirt HP, Belser UC, Lang NP. Long-term evaluation of non-submerged ITI implants. *Clin Oral Impl Res* 1997;8:161-172.
- Isidor F. Loss of osseointegration caused by occlusal load of oral implants: A clinical and radiographic study in monkeys. *Clin Oral Impl Res*

참 고 문 헌

- 1996;7:143-152.
17. Rangert B, Krogh PHJ, Langer B, Roekel NV. Bending overload and implant fracture: A retrospective clinical analysis. *Int J Oral Maxillofac Implants* 1995;10:326-334.
 18. Goodacre CJ, Kan JYK, Rungcharassaeng K. Clinical complications of osseointegrated implants. *J Prosthet Dent* 1999;81:537-552.
 19. Alsaadi G, Quirynen M, Koma'rek A, van Steenberghe D. Impact of local and systemic factors on the incidence of oral implant failures, up to abutment connection. *J Clin Periodontol* 2007;34:610-7.
 10. Moy PK, Medina D, Shetty V, Aghaloo TL. Dental implant failure rates and associated risk factors. *Int J Oral Maxillofac Implants* 2005;20:569-77.
 21. Chrcanovic BR, Kisch J, Albrektsson T, Wennerberg A. Factors influencing early dental implant failures. *J Dent Res* 2016;95:995-1002.
 22. Chung CH, Son MK. The classification and comparison of implant prosthesis according to types of retention. Part I: Screw retained prosthesis vs cement retained prosthesis. *J Korean Acad Oral Maxillofac Implantology* 2010;14:138-151.
 23. Hebel KS, Gajjar RC. Cement-retained versus screw-retained implant restoration: Achieving optimal occlusion and esthetics in implant dentistry. *J Prosthet Dent* 1997;77:28-35.
 24. Kallus T, Bessing C. Loose gold screws frequently occur in full arch fixed prostheses supported by osseointegrated implants after 5 years. *Int J Oral Maxillofac Implants* 1994;9:169-178.
 25. Carlson B, Carlsson GE. Prosthodontic complications in osseointegrated dental implant treatment. *Int J Oral Maxillofac Implants* 1994;9:90-94.
 26. Jemt T, Linden B, Lekholm U. Failures and complications in 127 consecutively placed fixed partial prostheses supported by Branemark implants: From prosthetic treatment to first annual checkup. *Int J Oral Maxillofac Implants* 1992;7:40-44.
 27. Park SD, Lee Y, Kim YL, Yu SH, Bae JM, Cho HW. Microleakage of different sealing materials in access holes of international connection implant systems. *J Prosthet Dent* 2012;108:173-180.
 28. Byrne D, Houston F, Cleary R, Claffey N. The fit of cast and premachined implant abutments. *J Prosthet Dent* 1998;80:184-192.
 29. Weng D, Jacobson Z, Tarnow D, Hurzeler MB, Faehn O, Sanavi F, Barkvoll P, Stach RM. A prospective multicenter clinical trial of 3i machined-surface implants: Results after 6 years of follow-up. *Int J Oral Maxillofac Implants* 2003;18:417-423.
 30. Herrmann I, Lekholm U, Holm S, Kultje C. Evaluation of patient and implant characteristics as potential prognostic factors for oral implant failure. *Int J Oral Maxillofac Implants* 2005;20:220-230.
 31. Henriksson K, Jemt T. Evaluation of custom-made procera ceramic abutments for single-implant tooth replacement: a prospective 1-year follow-up study. *Int J Prosthodont* 2003;16:626-30.
 32. Kim HS. Fabrication of custom abutment using dental CAD/CAM system. *J Korean Dent Assoc* 2012;50:118-25.
 33. Wu T, Liao W, Dai N, Tang C. Design of a custom angled abutment for dental implants using computer-aided design and nonlinear finite element analysis. *J Biomech* 2010;43:1941-6.
 34. Sailer I, Zembic A, Jung RE, Siegenthaler D, Holderegger C, Ha¨mmerle CH. Randomized controlled clinical trial of customized zirconia and titanium implant abutments for canine and posterior single-tooth implant reconstructions: preliminary results at 1 year of function. *Clin Oral Implants Res* 2009;20:219-25.
 35. Baldassarri M, Hjerpe J, Romeo D, Fickl S, Thompson VP, Stappert CF. Marginal accuracy of three implant-ceramic abutment configurations. *Int J Oral Maxillofac Implants* 2012;27:537-43.
 36. Lee CJ, Yang SE, Kim SG. Evaluation of reverse torque value of abutment screws on CAD/CAM custom-made implant abutments. *J Korean Acad Prosthodont* 2012;50:128-